

> Denotes revised material in this JPM

JOB PERFORMANCE MEASURE

TASK CODE: CEO-082

TASK: Restore Power to a CAM

NAME: _____ **>BADGE:** _____

>REFERENCES: Only references for knowledge items are listed here. The trainee is expected to identify the correct references for practical items.

- >1. WP 12-HP1300, Radiological Monitoring Equipment

TERMINAL OBJECTIVE:

Given that power has been secured to a CAM, restore power to the CAM per WP 12-HP1300.

CONSEQUENCES OF INADEQUATE PERFORMANCE:

Improper CAM operation

HAZARDS (PERSONNEL/EQUIPMENT STATUS):

Electrical Shock

PRE-REQUISITE TRAINING/ TASK COMPLETION:

- 1. CL 1.00 Series
- >2. CL 2.18 Air Sampling Equipment
- 3. ELC-103

TOOLS/EQUIPMENT (MATERIALS REQUIRED):

- | | |
|-------------------------|----------------------------------|
| 1. CAM | 3. Assorted slotted screwdrivers |
| 2. Key to a CAM Cabinet | >4. Log Sheets |

> Denotes revised material in this JPM

Instructions to Trainee: You shall acquire the necessary references and equipment, and complete all required documentation. Knowledge requirements shall be completed with 80% or greater accuracy. Critical step performance shall be completed with 100% accuracy.

Instructions to JPM Evaluator: The trainee is to perform the terminal objective, without assistance, on the job site. Provide clarification of requirements if requested by the trainee. You are encouraged to ask relevant questions to verify trainee understanding. If the trainee fails this JPM, clearly document the reason for failure and forward to the trainee's manager. Successful completion of this JPM shall be recorded on the trainee's qualification card.

KNOWLEDGE REQUIREMENTS:

Reference	Knowledge Requirement	Pass/Fail
>1	Discuss the purpose of the backup power supply to a CAM.	
1	Discuss the precautions associated with HV Power Switch while restoring power to a CAM	
1	State the minimum warm up time associated with the CAM while restoring power.	
1	Describe how the sequence of restoring power may vary for the different CAMs being utilized at WIPP	
1	State how to determine if a CAM has a UPS associated with it.	
1	State who must be notified after restoring power to a CAM	

PERFORMANCE REQUIREMENTS:

>Method	Performance Requirement	Pass/Fail
P, S	Restore power to an RMS Skid. #	
P, S	Restore power to an EMS Skid. #	
P, S	Restore power to an Auxiliary Data Acquisition Personal Computer. #	
P, S	Perform the required operability checks after power has been restored. #	
P, S	Complete the required documentation. #	

indicates a critical step

> Denotes revised material in this JPM

FINAL EVALUATION:

PASS

FAIL

COMMENTS:

EVALUATOR SIGNATURE:

_____ **DATE:** _____

TRAINEE SIGNATURE:

_____ **DATE:** _____

MANAGER SIGNATURE:

_____ **DATE:** _____